Certificate

- Graduation: 0.01mm
- With limit hands
- Rang per revolution: 1mm
- Supplied with lug and flat back

229-030-11

unit: mm



Styli(optional)

Application

Code	Range	Accuracy	Hysteresis	Scale	a	b	c	d	e	f
229-020-11	20mm	25μm	5μm	0-100	20	35	83	37.5	Ø8	Ø58
229-030-11	30mm	35μm	7μm	0-100	20	40	88	37.5	Ø8	Ø58

With manufacturer inspection certificate