228 DIAL INDICATOR

Certificate

Jeweled

- Bezel: Ø40mm
- Graduation: 0.001mm
- With limit hands
- Jeweled bearing
- Supplied with lug and flat back

Application



228-001-01

unit: mm



Contact points **(optional)**

Series **270**

Code	Range	Accuracy	Hysteresis	Range/rev	Scale
228-001-01	1mm	5μm	2μm	0.1mm	0-100

With manufacturer inspection certificate