

Code	Range	Accuracy	т	L	W	Remark
990-006-11	150mm/6"	±0.18mm	1mm	170mm	18mm	graduation on front and back
990-008-11	200mm/8"	±0.18mm	1mm	220mm	18mm	graduation on front and back
990-012-11	300mm/12"	±0.27mm	1mm	330mm	25mm	graduation on front and back
990-020-11	500mm/20"	±0.27mm	1mm	530mm	30mm	graduation on front and back
990-040-11	1000mm/40"	±0.45mm	1.5mm	1040mm	32mm	graduation on front and back
990-060-11	1500mm/60"	±0.63mm	1.5mm	1550mm	38mm	graduation on front
990-080-11	2000mm/80"	±0.80mm	2mm	2065mm	38mm	graduation on front