SERIES 222 DIAL INDICATOR

Certificate

Jeweled

- Graduation: 0.01mm
- With limit hands
- Jeweled bearing
- Range per revolution: 1mm
- With lock screw

222-010-11

Styli**(optional)**

Series 270

Application



Code	Range	Back	Accuracy	Hysteresis	Scale
222-003-11	3mm	lug	14μm	3μm	0-100
222-005-11	5mm	lug	16μm	3μm	0-100
222-010-11	10mm	lug	17μm	3μm	0-100
222-003-12	3mm	flat	14μm	3μm	0-100
222-005-12	5mm	flat	16μm	3μm	0-100
222-010-12	10mm	flat	17μm	3μm	0-100

With manufacturer inspection certificate