SERIES 221 DIAL INDICATOR

Certificate

Jeweled

- Graduation: 0.001mm
- With limit hands
- Jeweled bearing
- Rang per revolution: 0.2mm
- Supplied with lug and flat back

221-001-01

unit: mm



Styli**(optional)**

Application

Code	Range	Scale	Accuracy	Hysteresis	a	b	c	d
221-001-01	1mm	0-200	5μm	2μm	15	19.5	62.5	45
221-003-01	3mm	0-200	10μm	4μm	10	21.5	59.5	41.5
221-005-01	5mm	0-200	10μm	4μm	10	21.5	59.5	41.5

With manufacturer inspection certificate